



Certificate

Certificate no.

CU 72063355 01

License Holder:

Ergotron Inc.
1181 Trapp Road
St. Paul MN 55121
USA

Manufacturing Plant:

Ergotron Inc.
1181 Trapp Road
St. Paul MN 55121
USA

Test report no.: USA-ZZ 30582647 004

Client Reference: Anton Krycuk

Tested to: UL 60601-1:2003 R6.03
CAN/CSA-C22.2 No. 601.1-M90
UL 1778:2003 R3.03

Certified Product: StyleView Power System

License Fee - Units

Model Designation: SV22-22001, SV22-22002, SV22-22009

7

Rated Voltage: AC 120V, 60Hz
Rated Power: 5.1A
Protection Class: I
Output Ratings: AC 120V, 60Hz; 400VA, 300W

Special Remarks: To be used according to the user's manual and instructions specified by the licensee.
(UL 1778 considered for the UPS.)

7

Appendix: 1, 1-2

Inh. = 750351 / Deb. = 750351 / Fert. = 750351

Licensed Test mark:



Signatures

Date of Issue

(day/mo/yr)

15/01/2007

Stephan Schmitt
President

Dipl.-Ing. M. Raap
QA Certification Officer